

# EC Type-Examination Certificate

## 1572 Issue 7 Extension 7

This is to certify that the product group reference "Nitrile Powder-free" comprising the following products:

Product Ref.	Description		
Nitrile Powder-free	Powder-free disposable nitrile examination gloves. Available in the following range of colours:		
MaiMed-solution PF MaiMed-solution 100 PF MaiMed-Solution next PF	Powder-free: Blue, White,		
<b>Classification</b>			
EN 388:2003	Level EN374-3:2003	Level	
Abrasion resistance	0	Formaldehyde (15%)	6
Blade cut resistance	0	Sodium Hydroxide (40%)	6
Tear resistance	0	n-heptane	0
Puncture resistance	0	Sulphuric Acid (47%)	6

### Technical reports:

SATRA: 26448/0151/SPC-0, 0155685/0736/BT, SPC0215820/1326/SMcD/A  
TUV:7191004699-CHM11-01-TSL,7191004699-CHM11-02-TSL,  
7191004699-CHM11-03-TSL.


has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the complex category through:

- i. Testing to the following standard: EN 374-1:2003 Excluding cause 5.3.2
- ii. Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

1. Full details of the certification and product are contained in the manufacturer's technical file
2. This certificate is only valid if embossed with the text – "SATRA – European Notified Body 0321"
3. This certificate is issued subject to the conditions on the reverse side of this certificate
4. CE marking of production is also reliant on current compliance with Directive 89/686/EEC Article 11

Signed:  (D Cotter )

Signed: 

On behalf of SATRA

Date: 8<sup>th</sup> January 2015.